

Certificate of Analysis

Product Description: Material Number: Batch:	HBc IgM Positive Greater Than 4 S/CO Plasma 0305-0017, 0305-0018, 0305-0019, 0305-0020 9220874
Donor ID:	BD200258
Gender:	NA
Age:	NA
Country of Origin:	NA
Collection Date:	19 JUL 2007
Anticoagulant:	NA
Storage:	-20 degrees C
Shipping Conditions:	Dry ice
Intended Use:	For research use or for further manufacture into in-vitro diagnostic reagents.
Safety Precautions:	Use universal precautions when handling this material.
Test Method(s):	See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis							
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes		
24 Feb 2022	HBc IgM	Abbott Architect	14.470	S/CO	NA		
17 Oct 2017	HBc IgM	DiaSorin EIA Plus	9.237	S/CO	Neat		
17 Oct 2017	HBc IgM	DiaSorin EIA Plus	7.716	S/CO	1:2 dilution		
17 Oct 2017	HBc IgM	DiaSorin EIA Plus	5.397	S/CO	1:4 dilution		
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.							

Additional Data								
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes			
17 Feb 2022	HBeAg	DiaSorin Liaison XL	0.351	INDEX	≥0.9 = POS			
04 Oct 2017	HBs Antibody (anti-HBs)	Abbott Architect	>1000	mIU/mL	NA			
24 Feb 2022	HBsAg	Abbott Architect	0.210	S/CO	NA			

NA = Not Available

Approval:

Lynne Laborte

Prepared By

Dranni Parker

QA Verified By

28 MAR 2022

Date

29 MAR 2022

Date