

# Certificate of Analysis

Product Description: HBc Total Positive, HBsAg Negative Plasma  
 Material Number: 0305-0029, 0305-0030, 0305-0031, 0305-0032  
 Batch: 10322439  
 Donor ID: BD355407  
 Gender: NA  
 Age: NA  
 Country of Origin: United States  
 Collection Date: 13 MAY 2014  
 Anticoagulant: NA  
 Storage: -20 Degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
14-Mar-18	HBc Total	Abbott Architect	9.580	S/CO	NA
This product was tested for HBsAg, Anti-HCV, and anti-HIV 1/2 and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**



25 MAR 2019

Prepared By

Date



26 MAR 2019

QA Verified By

Date