

Certificate of Analysis

Product Description: HBeAg Positive Greater Than 2 S/CO Plasma
Material No(s): 0305-0053, 0305-0054, 0305-0055, 0305-0056
Lot No: BM201562
Donor ID: BD100781
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: 30-May-02
Anticoagulant: NA
Storage Temperature: - 20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis				
Test Date	Analyte	Test Method/Manufacturer	Result	Units
15-Aug-11	HBeAg	DiaSorin EIA	47.619	S/CO
12-Aug-11	HBsAG	Bio-Rad/ genetic Systems EIA	32.609	S/CO

This product was tested for a-HIV1/2 and a-HCV, and found to be negative.

Data are provided for informational purposes.
SeraCare Life Sciences does not claim that others can duplicate test results exactly.
NA = Not Available

Approval:



27 FEB 15

QA Prepared By

Date



02 MAR 15

QA Verified By

Date

