

# Certificate of Analysis

Product Description: HBsAg Positive Plasma (Quantitative)  
 Material Number: 0305-0061, 0305-0062, 0305-0063, 0305-0064  
 Batch: 9219554  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: 22 SEP 2006  
 Anticoagulant: NA  
 Storage: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

**Data are provided for informational purposes.**

**LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.**

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18 Aug 2022	HBsAg	Abbott Architect	1,462.260	S/CO	NA
27 Jan 2023	HBsAg	Abbott Architect Europe	90.940	IU/ML	≥0.05 IU/ML is Pos
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
19 Aug 2022	Hepatitis Delta IgG	DiaSorin Liaison XL	<0.100	AU/ML	NA

NA = Not Available

**Approval:**

*Lynne Labonte*

03 FEB 2023

Prepared By

Date

*Gaula J. Kumar*

08 FEB 2023

QA Verified By

Date