

# Certificate of Analysis

Product Description: HCV RNA Positive, HCV Antibody Negative Plasma  
 Material Number: 0310-0036, 0310-0037, 0310-0038  
 Batch: BM206306  
 Donor ID: BD111332  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: 26 JUL 2006  
 Anticoagulant: NA  
 Storage: -70 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

**Data are provided for informational purposes.**  
**LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.**

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
15 Dec 2023	HCV RNA	Roche cobas 6800	376,000.000	IU/ML	NA
This product was tested for Antibodies to HIV 1/2, HBsAg, HIV 1 RNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
05 Dec 2023	HCV Ab	Abbott Architect	0.100	S/CO	NA

NA = Not Available

**Approval:**

*Lynne Labonte*

18 DEC 2023

Prepared By

Date

*Gaula J. Keman*

20 DEC 2023

QA Verified By

Date