

# Certificate of Analysis

Product Description: HCV RNA TYPE 2B Plasma  
 Material No(s): 0310-0054, 0310-0055, 0310-0056  
 Lot No: 9207625  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: NA  
 Anticoagulant: NA  
 Storage Temperature: -70°C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
05-Apr-07	HCV Genotype	Siemens NS5B Genotyping	2B	NA	NA
01-Nov-15	HCV RNA	Roche COBAS AmpliPrep/TaqMan 2.0	191,837	IU/ML	NA
This product was tested for antibody to HIV-1/2 and HBsAg and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

*Brenda de Freitas*

28 Dec 15

Prepared By

Date

*Tsering Cheloni*

18 Jan 16

QA Verified By

Date