

Certificate of Analysis

Product Description: HCV RNA Type 2b Plasma
Material No(s): 0310-0054, 0310-0055, 0310-0056
Lot No: 9255088
Donor ID: BD111153
Demographics: NA
Collection Date: 06-Sep-11
Anticoagulant: NA
Storage Temperature: -70 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below


Lot Analysis

Test Date	Analyte	Test Method/Manufacturer	Result	Units
28-Nov-11	HCV	Ortho EIA	0.01	S/CO
19-Dec-11	HCV Genotype	Reference lab. (Quest)	2b	NA
20-Apr-12	HCV RNA	Roche Cobas Ampliprep/Taqman	11400000	IU/mL

This product was tested for a-HIV and HBsAg and found to be negative.

Data are provided for informational purposes.
SeraCare Life Sciences does not claim that others can duplicate test results exactly.
NA = Not Available

Approval:


QA


Date

