

Certificate of Analysis

Product Description: HCV RNA Type 4 Plasma
 Material Number: 0310-0060, 0310-0061, 0310-0062
 Batch: 9219511
 Donor ID: BD105431
 Gender: Female
 Age: 43
 Country of Origin: N/A
 Collection Date: 14 SEP 2007
 Anticoagulant: NA
 Storage: -70 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
01-Feb-19	HCV Genotype	Siemens/Bayer LIPA 2.0 Genotyping	4ACD	NA	NA
04-Feb-19	HCV RNA	Roche COBAS AmpliPrep/TaqMan 2.0	22,303.000	IU/ML	NA
04-Feb-19	HCV RNA	Roche COBAS AmpliPrep/TaqMan 2.0	21,430.000	IU/ML	NA
04-Feb-19	HCV RNA	Roche COBAS AmpliPrep/TaqMan 2.0	26,563.000	IU/ML	NA
14-Nov-07	HCV Ab	Ortho EIA	0.344	S/CO	NA
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, Syphilis and found to be negative.					

NA = Not Available

Approval:



09 FEB 2021

Prepared By

Date



11 FEB 2021

QA Verified By

Date