

Certificate of Analysis

Product Description: HIV-1 RNA Positive, HIV 1/2 AB Positive Plasma
Material No(s): 0315-0039, 0315-0040, 0315-0041
Lot No: 9122649
Donor ID: NA
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: NA
Anticoagulant: NA
Storage Temperature: -70 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
03-Nov-14	HIV-1 RNA	Roche COBAS AmpliPrep/TaqMan 2.0	23224	C/mL	
25-Aug-09	HIV-1 Antibody	Genetic Systems EIA	5.842	S/CO	
24-Aug-09	HIV-1 Antibody	Bio-Rad Rapid Test (Multispot)	Positive	NA	
24-Aug-09	HIV-2 Antibody	Bio-Rad Rapid Test (Multispot)	Negative	NA	
This product was tested for Antibody to HCV, HBsAg and Syphilis and found to be negative.					
Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
19-Aug-09	HIV-1 Antibody	MedMira Reveal Rapid Test	Positive	NA	
24-Aug-09	HIV-1 Antibody	Trinity Biotech Rapid Test	Positive	NA	
25-Aug-09	HIV-1/2 Antibody	Chembio Stat-Pak Rapid Test	Positive	NA	
24-Aug-09	HIV-1/2 Antibody	OraQuick Rapid Test	Positive	NA	
21-Aug-09	HIV-1 p24 Ag	Perkin Elmer EIA	0.633	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

05-Aug-15

Prepared By

Date

Christine Pawia

07-Aug-15

QA Verified By

Date

