

Certificate of Analysis

Product Description: CMV IgG Positive Plasma
Material No(s): 0320-0015, 0320-0016, 0320-0017, 0320-0018
Lot No: BM203102
Donor ID: BD101662
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: 07-Nov-05
Anticoagulant: NA
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Nov-11	CMV IgG	Siemens Immulite	8.85	Index	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

01-Oct-15

Prepared By

[Signature]

Date

06-Oct-15

QA Verified By

Date

