

Certificate of Analysis

Product Description: Rubella IgM Positive Plasma
 Material Number: 0320-0027, 0320-0028, 0320-0029, 0320-0030
 Batch: 10641558
 Donor ID: BD509028
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 18 JUL 2022
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 Degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

| Batch Analysis | | | | | |
|----------------|-------------|--------------------------|---------|-------|----------------|
| Test Date | Analyte | Test Method/Manufacturer | Result | Units | Notes |
| 14 Oct 2022 | Rubella IgM | DiaSorin Liaison XL | 113.000 | AU/ML | Neat |
| 14 Oct 2022 | Rubella IgM | DiaSorin Liaison XL | 56.500 | AU/ML | 1:1.5 dilution |

This product was tested for HBsAg, Antibodies to HIV 1/2, Antibodies to HTLV I/II, Antibodies to HCV and Syphilis and found to be negative.

NA = Not Available

Approval:



29 NOV 2022

Prepared By

Date



02 DEC 2022

QA Verified By

Date