

Certificate of Analysis

Revised 31JAN2022

Product Description: Rubella IgM Positive Plasma
 Material Number: 0320-0027, 0320-0028, 0320-0029, 0320-0030
 Batch: 10363709
 Donor ID: BD328432
 Gender: NA
 Age: NA
 Country of Origin: US
 Collection Date: 11 APR 2017
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
01 Aug 2018	Rubella IgM	Bio-Rad EIA	7.220	INDEX	Neat
01 Aug 2018	Rubella IgM	Bio-Rad EIA	3.140	INDEX	1:4 dilution
25 Jul 2019	Rubella IgM	DiaSorin Liaison	142.000	AU/ML*	Neat
25 Jul 2019	Rubella IgM	DiaSorin Liaison	159.000	AU/ML*	Neat
25 Jul 2019	Rubella IgM	DiaSorin Liaison	151.000	AU/ML*	Neat
19 Sep 2018	Rubella IgM	DiaSorin Liaison	132.000	AU/ML*	Neat
25 Jul 2019	Rubella IgM	DiaSorin Liaison	32.200	AU/ML*	1:4 dilution
25 Jul 2019	Rubella IgM	DiaSorin Liaison	27.900	AU/ML*	1:4 dilution
25 Jul 2019	Rubella IgM	DiaSorin Liaison	26.900	AU/ML*	1:4 dilution
05 Aug 2019	Rubella IgM	Roche cobas 6000	3.280	COI	Neat
05 Aug 2019	Rubella IgM	Roche cobas 6000	3.300	COI	Neat
05 Aug 2019	Rubella IgM	Roche cobas 6000	3.290	COI	Neat
05 Aug 2019	Rubella IgM	Roche cobas 6000	1.070	COI	1:4 dilution
05 Aug 2019	Rubella IgM	Roche cobas 6000	1.040	COI	1:4 dilution
05 Aug 2019	Rubella IgM	Roche cobas 6000	1.070	COI	1:4 dilution
05 Aug 2019	Rubella IgM	Siemens Centaur	4.190	INDEX	Neat
05 Aug 2019	Rubella IgM	Siemens Centaur	4.390	INDEX	Neat
05 Aug 2019	Rubella IgM	Siemens Centaur	4.070	INDEX	Neat
05 Aug 2019	Rubella IgM	Siemens Centaur	1.340	INDEX	1:4 dilution
05 Aug 2019	Rubella IgM	Siemens Centaur	1.180	INDEX	1:4 dilution
05 Aug 2019	Rubella IgM	Siemens Centaur	1.320	INDEX	1:4 dilution

This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.

*≥20 AU/ML is positive based on site utilized

NA = Not Available

Certificate of Analysis

Revised 31JAN2022

Product Description: Rubella IgM Positive Plasma
Material Number: 0320-0027, 0320-0028, 0320-0029, 0320-0030
Batch: 10363709

Approval:

Synne Labonte

14 JAN 2022

Prepared By

Date

Danni Parker

31 JAN 2022

QA Verified By

Date