

Certificate of Analysis

Product Description: Chagas AB Positive Plasma
 Material Number: 0325-0040, 0325-0041, 0325-0042
 Batch: 10653047
 Donor ID: BD512814
 Gender: Female
 Age: 23
 Country of Origin: Honduras
 Collection Date: 21 APR 2022
 Anticoagulant: CPDA-1
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.
LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
30 Jan 2023	Chagas	Ortho EIA	2.473	S/CO	Neat
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

NA = Not Available

Approval:



31 MAR 2023

Prepared By

Date



06 APR 2023

QA Verified By

Date