

Certificate of Analysis

Product Description: Zika IgG Positive Plasma
 Material Number: 0325-0072, 0325-0073, 0325-0074, 0325-0075
 Batch: 10226463
 Donor ID: BD309168
 Gender: Female
 Age: 44
 Country of Origin: Dominican Republic
 Collection Date: 18 FEB 2016
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 Degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
14 Oct 2016	Zika IgG	Euroimmun EIA	>200	RU/ML	>22 RU/ML is Positive
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, and HBsAg and found to be negative					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
25 Jan 2017	Zika IgM	InBios EIA	13.825	ISR	NA

NA = Not Available

Approval:



20 DEC 2021

Prepared By

Date



21 DEC 2021

QA Verified By

Date