

Certificate of Analysis

Product Description: ANA Speckled Positive Plasma
 Material No(s): 0340-0017, 0340-0018, 0340-0019, 0340-0020
 Lot No: 9215950
 Donor ID: BD100155
 Gender: Male
 Age: NA
 Country of Origin: NA
 Collection Date: 24-Feb-07
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
11-Feb-10	Anti-Nuclear Antibody	Immunofluorescent Assay (Quest Diag.)	≥ 1:1,280 Speckled	Titer Pattern	
14-Aug-07	Anti-Nuclear Antibody	Diamedix EIA	6.043	S/CO	
This product was tested for HBsAg, Antibody to HIV1/2, Antibody to HCV, HIV1 RNA, HCV RNA, and HBV DNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
13-Aug-07	Smith Antibody	Diamedix EIA	5.88	INDEX	
14-Aug-07	dsDNA Antibody	Diamedix EIA	31.6	IU/mL	Equivocal

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Brenda de Freitas

09Jun15

Prepared By

Date

Anne Moore

11Jun15

QA Verified By

Date

