

Certificate of Analysis

Product Description: EBV VCA IgM Positive Plasma
 Material Number: 0345-0001, 0345-0003, 0345-0004, 0345-0005
 Batch: 10530447
 Donor ID: BD454263
 Gender: NA
 Age: NA
 Country of Origin: United States
 Collection Date: 17 NOV 2019
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
04-Dec-20	EBV VCA IgM	Zeus EIA	8.891	S/CO	NA
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

NA = Not Available

Approval:



14 JAN 2021

Prepared By

Date



18 JAN 2021

QA Verified By

Date