

Certificate of Analysis

Product Description: EBV VCA IgG Positive Plasma
 Material No(s): 0345-0006, 0345-0007, 0345-0008, 0345-0009
 Lot No: BM216620
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 18-Apr-03
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
09-Jan-07	EBV VCA IgG	Diamedix EIA	4.231	S/CO	NA
30-Jan-07	EBV VCA IgG	DiaSorin EIA	6.340	S/CO	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIVAg, HTLV I/II, Syphilis and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Nov-06	EBV EA IgG	Diamedix EIA	0.249	S/CO	NA
17-Jan-11	EBV EA IgG	Wampole EIA	0.197	S/CO	NA
19-Dec-06	EBV EA-D IgM	Diamedix EIA	0.792	S/CO	NA
20-Dec-06	EBV EBNA IgG	Trinity Biotech Captia EIA	2.410	S/CO	NA
23-Apr-07	EBV EBNA IgM	Diamedix EIA	0.312	S/CO	NA
19-Dec-06	EBV VCA IgM	Diamedix EIA	0.207	S/CO	NA
24-Apr-07	EBV VCA IgM	Wampole EIA	0.036	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabente

23-Nov-15

Prepared By

Date

Tsering Chelung

02-Dec-15

QA Verified By

Date