

Certificate of Analysis

Product Description: EBV IgG Negative Plasma
 Material No(s): 0345-0010, 0345-0011, 0345-0012, 0345-0013
 Lot No: 10127774
 Donor ID: BD111442
 Gender: Male
 Age: 25
 Country of Origin: US
 Collection Date: 26-Jun-15
 Anticoagulant: 4% Sodium Citrate
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
27-Jul-15	EBV EA IgG	Zeus EIA	0.424	S/CO	NA
27-Jul-15	EBV EBNA IgG	Zeus EIA	0.203	S/CO	NA
27-Jul-15	EBV VCA IgG	Wampole EIA	0.097	S/CO	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA, Syphilis and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

20-Nov-15

Prepared By

Date

Tsering Chodoyi

04-Dec-15

QA Verified By

Date