

# Certificate of Analysis

Product Description: Syphilis (T. Pallidum) IgM Positive Plasma  
 Material Number: 0365-0009, 0365-0010, 0365-0011, 0365-0012  
 Batch: 9195739  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: NA  
 Anticoagulant: NA  
 Storage: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

**Data are provided for informational purposes.**  
**LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.**

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
22 Jan 2016	Syphilis IgM	Calbiotech EIA	5.792	S/CO	OFF SCALE
This product was tested for Antibodies to HIV1/2, Antibodies to HCV, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	Neat
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:2 dilution
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:4 dilution
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:8 dilution
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:16 dilution
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:32 dilution
30 Oct 2023	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	NEG	NA	1:64 dilution

NA = Not Available

**Approval:**

*Sydney Rabente*

01 NOV 2023

Prepared By

Date

*Gaula J. Kuman*

09 NOV 2023

QA Verified By

Date