

Certificate of Analysis(Revised)

Research Use Only - Not for Use in Diagnostic Procedures

Product Description:	Product	Material Code	Batch Number
	AccuPlex™ SARS-CoV-2 Verification Panel - Full Genome	0505-0168	10499611

Date of Manufacture: 16 JUN 2020

Expiration Date: 16 JUN 2022

Storage Temperature: 2 - 8 °C

Viral copies/mL for Member 1 determined using TaqMan® digital PCR Assays run on the BioRad QX200 system.

Kit Component	Quantitative Result (cp/mL)
SARS-CoV-2 Member 1	88,000 cp/mL

QC Testing:

Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for Disease Control Assay publication and instructions for use¹.

Kit Component	Qualitative Result
SARS-CoV-2 Member 1	Positive
SARS-CoV-2 Member 2	Positive
SARS-CoV-2 Member 3	Positive
SARS-CoV-2 Negative	Negative

Safety Precautions: Preparation contains recombinant Sindbis virus. Use the Centers for Disease Control recommended universal precautions.

CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use. CDC-006-00019 Revision: 03. Effective 330/2020.

Approval:

<p><i>Catherine Huang</i> Prepared By</p>	<p>25 JAN 2021 Date</p>
<p><i>Kimberly M Waller</i> QA Verified By</p>	<p>25 JAN 2021 Date</p>