

Certificate of Analysis

Product Description: AccuPlex™ SARS-CoV-2 in Synthetic Oral Fluid

Material Number: 0505-0222

Batch Number 10531204

Date of Manufacture 03 Dec 2020

Expiration Date 12 Jan 2023

Storage Temperature 2-8°C

Viral copies/mL for Member 1 determined using TagMan® digital PCR Assays run

on the BioRad QX200 system

Kit Component	SARS-CoV-2 Nucleocapsid region (cp/mL)
Positive Member 1	2.06E+04

¹Real time qPCR testing using CDC nCoV 2019 primers/ and probes described in the SalivaDirect™ protocol

QC Testing

Kit Component	SalivaDirect [™] RNA Extraction-Free SARS-CoV-2 Testing
Positive Member 1	Positive
Positive Member 2	Positive
Positive Member 3	Positive
Negative Member 4	Undetermined

¹ SalivaDirect™: RNA extraction-free SARS-CoV-2 diagnostics

Grubaugh

https://www.protocols.io/view/salivadirect-rna-extraction-free-sars-cov-2-diagno-bkjgkujw

Safety Precautions

Preparation contains recombinant Alphavirus. Use the Centers for Disease Control

recommended universal precautions.

Approval:

Prepared By Date

Mindey / 11 Wodlood 22 JAN 3021

QA verified by Date