

# Revised Certificate of Analysis

## Research Use Only - Not for Use in Diagnostic Procedures

Product	Material Code	Batch Number
AccuPlex™ Flu A/B and RSV Verification Panel	0515-0002	10521190

**Date of Manufacture:** 29 SEP 2020

**Expiration Date:** 07 OCT 2023

**Storage Temperature:** 2 - 8 °C

Average viral copies/mL for Member 1 determined using TaqMan® digital PCR Assays run on the BioRad QX200 system.

Kit Component	Analyte	Quantitative Result (cp/mL)
Member 1	Flu A	1.17E+05
	Flu B	8.66E+04
	RSV	8.65E+04

Qualitative detection using Cepheid Xpert® Xpress Flu/RSV Assay run on the Cepheid GeneXpert instrument.

**QC Testing:**

Kit Component	Result
Member 1	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 2	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 3	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Negative	Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE

**Safety Precautions:**

Preparation contains recombinant Alphavirus. Use the Centers for Disease Control recommended universal precautions.

**Approval:**

Prepared By



QA Verified By

Date

27 OCT 2022

28 OCT 2022

Date