

# Certificate of Analysis

## Research Use Only - Not for Use in Diagnostic Procedures

	Product	Material Code	Batch Number
Product Description:	AccuPlex™ Flu A/B and RSV Verification Panel	0515-0002	10656762

Date of Manufacture: 14-FEB-2023

Expiration Date: 03-APR-2025

Storage Temperature: 2 - 8 °C

Average viral copies/mL for Member 1 determined using TaqMan® digital PCR Assays run on the BioRad QX200 system.

Kit Component	Analyte	Quantitative Result (cp/mL)
Member 1	Flu A	8.99E+05
	Flu B	1.15E+06
	RSV	9.16E+05

Qualitative detection using Cepheid Xpert® Xpress Flu/RSV Assay run on the Cepheid GeneXpert instrument.

QC Testing:

Kit Component	Result
Member 1	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 2	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 3	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Negative	Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE

Safety Precautions:

Preparation contains recombinant Alphavirus. Use the Centers for Disease Control recommended universal precautions.

Approval:



24-APR-2023

Prepared By

Date



25-APR-2023

QA Verified By

Date