

Certificate of Analysis

For Research Use Only
Not for Use in Diagnostic Procedures

Product Description: Seraseq™ Tumor Mutation DNA Mix v2 AF10, RUO
 Material No: 0710-0094 Batch No: 10656723
 Volume: 25 µL Concentration: 25 ng/µL
 Date of Manufacture: 16 FEB 2023 Expiration Date: 17 OCT 2024
 Test Method for Concentration: Qubit dsDNA BR Assay
 Average Concentration (ng/µL): 25.5 ng/µL
 Test Method for Variant Allele Frequency: In house validated, allele specific, digital PCR assays

Mutation	Average dPCR Allele Frequency
AKT1 COSM33765	9.8
APC COSM13127	10.1
APC COSM18561	10.0
ATM COSM21924	10.7
BRAF COSM476	10.2
CTNNB COSM5664	11.4
EGFR COSM6224	10.2
EGFR COSM6225	10.5
EGFR COSM6240	10.0
EGFR COSM12378	10.0
ERBB2 COSM682	9.7
FGFR3 COSM715	11.1
FLT3 COSM783	10.8
FOXL2 COSM33661	11.9
GNA11 COSM52969	10.4
GNAQ COSM28758	11.0
GNAS COSM27887	10.0
IDH1 COSM28747	9.8
JAK2 COSM12600	10.2
KIT COSM1314	9.8

Mutation	Average dPCR Allele Frequency
KRAS COSM521	9.8
MPL COSM18918	10.3
NPM1 COSM17559	9.9
NRAS COSM584	10.6
PDGFRA COSM736	10.2
PDGFRA COSM28053	9.6
PIK3CA COSM763	9.6
PIK3CA COSM775	10.9
PIK3CA COSM12464	10.1
PTEN COSM4986	10.1
PTEN COSM5809	10.1
RET COSM965	11.6
SMAD4 COSM14105	11.2
TP53 COSM6530	9.6
TP53 COSM10660	10.0
TP53 COSM10662	9.9
TP53 COSM10648	9.8
TP53 COSM18610	Not Assayed
NCOA4-RET	11.7
TPR-ALK	11.0

Approval:



17 MAR 2023

Prepared By



Date

30 MAR 2023

QA Verified By

Date