

# Certificate of Analysis

Product Description: ACCURUN Swab RSV Reference Material  
 Material Number: 2020-0179  
 Batch Number: 10684945  
 Storage Temperature: Ambient  
 Date of Manufacture: 16 NOV 2023  
 Date of Expiration: 16 NOV 2025

Testing was performed using Cepheid Xpert® Xpress CoV-2/Flu/RSV plus assay run on the GeneXpert system and ID NOW™ RSV assay performed on the ID NOW Instrument.

QC Testing:

Component	Test Method	Result
ACCURUN Swab RSV Reference Material	Cepheid	RSV Positive 31.42 Ct
	Abbott ID NOW	RSV Positive

Safety Precautions: Preparation contains heat inactivated RSV A2 virus. Use the Center for Disease Control recommended universal precautions.

**Approval:**

*Ravathi Raman*  
 Prepared By \_\_\_\_\_ Date 12 DEC 2023

*[Signature]*  
 QA Verified By \_\_\_\_\_ Date 13 DEC 2023