



CERTIFICATE OF ANALYSIS

BacTrace[®] E. coli O157:H7 Positive Control

CATALOG NO. 50-95-90
LOT NO. 10174333
SIZE 1.0 mL

DESCRIPTION Heat-killed *Escherichia coli* serotype O157:H7 cells (1 mL of approximately 2% wet packed cell w/v) in dextran solution.

LOT SPECIFIC CELL COUNT Cell count = 3.98×10^9 cfu/mL

FORM/STORAGE Lyophilized. Store refrigerated at 2 – 8°C.

SPECIFICATION Verification of heat kill is performed by incubating streaked TSA plates for 24 hours at 37° C:
No Growth = Pass

STABILIZERS AND PRESERVATIVES Dextran added to stabilize during lyophilization. No preservatives. Non-sterile.

STABILITY Product is stable for one year from date of receipt when stored properly.

DATE OF MANUFACTURE 03/30/2016

Quality Assurance Verified and Approved:

Date: 04/08/2016