

Certificate of Analysis

Product Description: RPR Positive Plasma
 Material No(s): 0365-0005, 0365-0006, 0365-0007, 0365-0008
 Lot No: 9242969
 Donor ID: BD110077
 Gender: Female
 Age: 19
 Country of Origin: NA
 Collection Date: 15-May-09
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	Neat
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:2 dilution
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:4 dilution
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:8 dilution
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:16 dilution
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	NEG	NA	1:32 dilution
08-Jan-10	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	NEG	NA	1:64 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, and HIV 1 RNA, HCV RNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
15-Dec-09	Syphilis IgG	Trinity Biotech Captia EIA	1.723	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

18-Dec-15

Prepared By

Date

Tsering Chelung

22-Dec-15

QA Verified By

Date