Certificate of Analysis

Rubella IgG Positive Plasma
0320-0023, 0320-0024, 0320-0025, 0320-0026
BM201626
BD103226
NA
NA
NA
22-Feb-02
4% Sodium Citrate
-20 degrees C
Dry ice
For research use only. Not for use in diagnostic procedures.
Use universal precautions when handling this material.
See below

Lot Analysis								
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes			
20-Sep-05	Rubella IgG	Diamedix EIA Quantitative	2.94	Index	29.4 IU/mL			
08-Dec-14	Rubella IgG	Trinity Biotech Captia EIA	4.031	s/co	(1:4 dilution) NA			
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HIVAg, HTLV I/II, Syphilis and found to be negative.								

Additional Data									
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes				
12-Sep-05	Rubella IgM	Diamedix EIA	0.239	S/CO	NA				

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly. NA = Not Available

Approval:

Lynne Laborte

Prepared By

QA Verified By

Date 07-Oct-15

Date

05-Oct-15



SeraCare.com

Milford, MA 01757 508.634.3334 Fax

508.244.6400