Certificate of Analysis

SSA AB Positive Plasma
0340-0081, 0340-0082, 0340-0083, 0340-0084
BM204566
BD101726
NA
NA
NA
09-Dec-05
NA
-20 degrees C
Dry ice
For research use only. Not for use in diagnostic procedures.
Use universal precautions when handling this material.
See below

Lot Analysis									
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes				
24-Feb-06	Sjögren's Syndrome A	Diamedix EIA	8.26	Index					
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV-1 RNA, HCV RNA, HBV DNA and found to be negative.									

Additional Data								
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes			
24-Feb-06	Anti-JO-1	Diamedix EIA	0.12	Index				
10-Feb-06	Anti-Nuclear Antibody	Diamedix EIA	3.15	S/CO				
24-Feb-06	Anti-SCL 70	Diamedix EIA	0.11	Index				
24-Feb-06	Anti-SM/RNP	Diamedix EIA	0.13	Index				
24-Feb-06	Anti-Smith	Diamedix EIA	0.04	Index				
24-Feb-06	Sjögren's Syndrome B	Diamedix EIA	0.08	Index				

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly. NA = Not Available

Approval:

Lynne Laborte Prepared By

Prepared By

Christine Paisa

QA Verified By



11-Aug-15

Date

21-Aug-15

Date

Milford, MA 01757 508.634.3334 Fax

508.244.6400

SeraCare.com