

AccuPlex™ SARS-CoV-2, Flu A/B and RSV Verification Panel

About this package insert

Thank you for your interest in this AccuPlex™ product. This package insert consists of two pages.

The first page contains the product name, the SeraCare logo, and contact information.

The second page contains the complete package insert text. If the package insert you view or print does not contain two pages, or if you experience any problems, please email us at info@seracare.com, or call us at +1.508.244.6400.

A printed package insert will be sent to you upon request.



SeraCare Life Sciences, Inc. | 37 Birch Street, Milford, MA 01757 USA
Phone: +1 508.244.6400 | info@seracare.com

13799CA-01 March 2021

AccuPlex™ SARS-CoV-2, Flu A/B and RSV Verification Panel

NAME AND INTENDED USE

AccuPlex™ SARS-CoV-2, Flu A/B and RSV Verification Panel (0505-0183) is formulated for use with multiplex test methods that can detect SARS-CoV-2, influenza A/B and respiratory syncytial virus. The panel contains different concentrations of positive reference material to enable evaluation of test performance at multiple points across the assay range; it is designed to support SARS-CoV-2, Flu A/B and RSV multiplex test verification at installation. AccuPlex virus products are non-replicative recombinant viruses that are intended to assess the performance of the full process of a molecular test. AccuPlex can be used to evaluate test proficiency and accuracy through the full process because they are encapsulated viruses which require extraction and amplification. *For Research Use Only. Not for use in diagnostic procedures.*

PRODUCT DESCRIPTION

This product contains recombinant Alphavirus. There are 3 vials of positive reference material that contain recombinant virus particles with following sequence coverage:

Virus	Genbank Accession Number	Regions Included
Influenza A	KU933490 - KU933497	Full Genome
Influenza B	CY236601.1- CY236608.1	Full Genome
RSV	NC_001803	1..4380; 8460..15191
SARS-CoV-2	NC_045512.2	Full Genome

Positive vials are included at the following concentrations:

Member 1 – 100,000 copies/mL, Member 2 – 10,000 copies/mL,
Member 3 – 1000 copies/mL.

There is also 1 vial of negative reference material that contains recombinant virus particles with sequences from human RNase P gene (RP) at a concentration of 5000 copies/mL.

The recombinant viruses used to produce the AccuPlex SARS-CoV-2, Flu A/B and RSV reference material are replication defective and heat-treated. However, handle AccuPlex products and all human blood products as though they can transmit infectious agents.

The product is formulated in viral transport media that consists of Tris-buffered saline, with added glycerol, anti-microbial agents and human proteins. **This material must go through extraction, similar to the patient sample.**

Material Number: 0505-0183
 Positive: Member 1: 1 x 3.0 mL; 100,000 copies/mL
 Member 2: 1 x 3.0 mL; 10,000 copies/mL
 Member 3: 1 x 3.0 mL; 1000 copies/mL
 Negative: Negative: 1 x 3.0 mL; 5000 copies/mL (RNase P)

STORAGE INSTRUCTIONS

This product should be stored at 2 - 8 °C during regular use. It may also be initially stored at -20 °C, but subsequently maintain thawed material at 2 - 8 °C. Do not expose to multiple freeze thaw cycles. Each vial can be used multiple times up until the date of expiry.

INSTRUCTIONS FOR USE

Allow the product vial to come to room temperature before use. Mix by vortexing to ensure a homogeneous suspension. AccuPlex reference materials should be added to a test run using the same procedure provided by the manufacturer for unknown specimens. AccuPlex SARS-CoV-2, Flu A/B and RSV Verification Panel members must go through an extraction process prior to detection by PCR. Process the product according to the instructions for unknown samples provided by the test kit or the laboratory's standard operating procedures. AccuPlex verification panel must NOT be substituted for the positive and negative control reagents provided with the manufactured test kits.

INTERPRETATION OF RESULTS

Levels of reactivity for the AccuPlex SARS-CoV-2, Flu A/B and RSV materials may vary with different types of tests and different test kit lots, but positive reference materials are expected to give positive results, and negative reference materials are expected to give negative results. This product contains targeted formulations of 100,000, 10,000, and 1000 copies/mL for all viral targets for positive members and 5000 copies/mL for the negative member as measured using reverse transcription digital PCR. Note that the positive reference material may contain traces of RNase P and therefore generate a positive RNase P result due to the presence of a human plasma component in the product matrix; it is not designed or intended to be used as an RNase P reference material.

LIMITATIONS OF THE PROCEDURE

AccuPlex SARS-CoV-2, Flu A/B and RSV Verification Panel must not be substituted for the control reagents provided with manufactured test kits. Test procedures and interpretation of results provided by manufacturers of test kits must be followed closely. Deviations from procedures recommended by test kit manufacturers may produce unreliable results. AccuPlex materials are not calibrators and should not be used for assay calibration. Performance characteristics for AccuPlex SARS-CoV-2, Flu A/B and RSV Verification Panel have been established only for amplified nucleic acid tests for RNA. Adverse shipping and storage conditions or use of outdated product may produce erroneous results.

WARNINGS AND PRECAUTIONS

Use Centers for Disease Control (CDC) recommended universal precautions for handling AccuPlex SARS-CoV-2, Flu A/B and RSV materials and human specimens¹. Do not pipette by mouth; do not smoke, eat or drink in areas where specimens are being handled. Clean any spillage by immediately wiping with 0.5% sodium hypochlorite solution. Dispose of all specimens and materials used in testing as though they contain infectious agents.

EXPECTED RESULTS

Specific levels of reactivity will vary among different manufacturers' assays, different procedures, different lot numbers and different laboratories. Procedures for implementing a quality assurance program and monitoring test performance on a routine basis must be established by each individual laboratory. Each laboratory should establish its own range of acceptable values.

REFERENCES

1. Siegel JD, Rhinehart E, Jackson M, Chiarello L, and the Healthcare Infection Control Practices Advisory Committee. 2007 Guideline for Isolation Precautions: Preventing Transmission of Infectious Agents in Healthcare Settings.

For assistance, contact SeraCare Technical Support at +1 508.244.6400.