

AccuSet™ HCV Worldwide Performance Panel 0810-0230

INTENDED USE

The AccuSet™ HCV Worldwide Panel 0810-0230 is intended for use by diagnostic manufacturers, researchers, and clinical laboratories to develop, evaluate, or troubleshoot HCV test methods. Characterized samples and comprehensive data are provided for comparative analysis. For research use only. Not for use in diagnostic procedures.

PRODUCT DESCRIPTION

This product is a 22-member panel consisting of undiluted, naturally occurring plasma samples collected worldwide from individual donors positive for HCV infection of varying genotypes (1, 2, 3, 4, 5 and 6). Each sample represents a single collection event. No preservatives were added.

Material Number: 0810-0230, 1 vial per member
22 members, 1.0 mL per vial

STORAGE

Panel members should be stored frozen at -70 °C or colder to preserve HCV RNA. SeraCare Life Sciences recommends that the panel members be divided into smaller aliquots to avoid multiple freeze-thaw cycles, if appropriate. If turbidity or particulate matter is observed, the samples should be centrifuged in accordance with each test kit manufacturer's instructions for sample preparation.

INTERPRETATION OF RESULTS

The Data Sheet for AccuSet HCV Worldwide Panel 0810-0230 is available at www.seracare.com. The Data Sheet lists results for panel members generated using commercially-available screening, monitoring, and supplemental test methods. Tests were performed at SeraCare or at recognized reference laboratories (RL) by individuals who routinely use these procedures. Information regarding specific test methods is available on the data sheet. Data Sheets are updated when new data are available.

LIMITATIONS

AccuSet HCV Worldwide Panel 0810-0230 is offered for research use only. Not for use in diagnostic procedures. Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

PRECAUTIONS

These materials have not been treated and should be considered biohazardous. Follow Universal Precautions¹. All panel members were found positive for HCV RNA and several members were found positive for anti-HCV. The units that make up this panel were tested and found negative for anti-HIV and HBsAg. This does not ensure the absence of these or other human pathogens.

Do not pipette by mouth. Do not smoke, eat, or drink in areas where specimens are handled.

These materials should be disposed of in a manner that will inactivate pathogenic agents.

REFERENCES

1. Siegel JD, Rhinehart E, Jackson M, Chiarello L, and the Healthcare Infection Control Practices Advisory Committee, 2007 Guideline for Isolation Precautions: Preventing Transmission of Infectious Agents in Healthcare Settings.

For assistance, contact SeraCare at 508.244.6400.

The package insert for this panel can be found at www.seracare.com.

A printed copy of the package insert or data sheet may be requested by email at info@seracare.com or by phone at 508.244.6400.