

Product Report

Product Description: Seraseq™ ctDNA Mutation Mix v2 WT

Material No: 0710-0144

Batch No: 10267221

Material Description: Mixture of human genomic DNA from the reference cell line, GM24385, and synthetic DNA constructs

Date of Manufacture: 27Jun2017

Expiration Date: 27Jun2018

Concentration (Qubit dsDNA BR Assay): 10.9 ng/ μ L

Volume: 25 μ L

Storage: -20 °C

Digital PCR testing using BioRad QX200™ Droplet Digital™ PCR System:

Gene ID	COSMIC Identifier	AF% ¹
AKT1	COSM33765	0.03
APC	COSM13127	0.05
APC	COSM18561	0.00
ATM	COSM21924	0.00
BRAF	COSM476	0.00
CTNNB1	COSM5664	0.02
EGFR	COSM6224	0.02
EGFR	COSM12378	0.02
EGFR	COSM6225	0.00
EGFR	COSM6240	0.02
ERBB2	COSM682/20959	0.00
FGFR3	COSM715	0.00
FLT3	COSM783	0.00
FOXL2	COSM33661	0.00
GNA11	COSM52969	0.00
GNAQ	COSM28758	0.00
GNAS	COSM27887	0.06
IDH1	COSM28747	0.08
JAK2	COSM12600	0.00
KIT	COSM1314	0.01
KRAS	COSM521	0.02
MPL	COSM18918	0.00
NCOA4-RET	N/A	0.00
NPM1	COSM17559	0.00
NRAS/CSDE1	COSM584	0.09
PDGFRA	COSM736	0.03
PDGFRA	COSM28053	0.00
PIK3CA	COSM763	0.03
PIK3CA	COSM12464	0.00
PIK3CA	COSM775	0.10
PTEN	COSM4986	0.00
PTEN	COSM5809	0.00
RET	COSM965	0.01
SMAD4	COSM14105	0.01
TP53	COSM10648	0.05
TP53	COSM10660	0.03
TP53	COSM10662	0.07
TP53	COSM6530	0.00
TP53	COSM18610	Not Tested ²
TPR-ALK	N/A	0.01

¹ Variant allele frequencies > 0.00% for this wildtype negative control are within the expected range for stochastic positive dPCR reactions.

² NGS testing alone is used for QC of variant COSM18610.

Next Generation Sequencing Testing:

Product was tested using the Archer[®] Reveal ctDNA[™] 28 Kit run on an Illumina[®] MiSeq[™]. No variants were called above statistical background.

Approval:



Prepared By

14 Jul 2017

Date