

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description:	Seraseq® ctDNA Complete Reference Material WT		
Material Number:	0710-0674	Batch Number:	10593197
Material Description:	A ctDNA-like mixture of human genomic DNA from the reference cell line, GM24385		
Concentration (Qubit dsDNA BR Assay):	Nominal value: 25 ng/mL; Average measured value after extraction using Qiagen QIAamp Circulating Nucleic Acid Kit: 27.5 ng/mL		
Fill Volume:	5.0 mL		
Date of Manufacture:	22 NOV 2021	Expiration Date:	22 NOV 2024
Storage:	2-8°C		

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material WT
 Material Number: 0710-0674 Batch Number: 10593197

Digital PCR testing using
 BioRad QX200™ Droplet
 Digital™ PCR System:

Gene ID	COSMIC Identifier	Amino Acid Change	Average AF% ¹
AKT1	COSM33765	p.E17K	0.005
BRAF	COSM476	p.V600E	0.000
EGFR	COSM6224	p.L858R	0.001
EGFR	COSM6240	p.T790M	0.009
ERBB2	COSM20959	p.A775_G776insYVMA	0.007
KIT	COSM1314	p.D816V	0.002
KRAS	COSM521	p.G12D	0.012
NCOA4/RET	NA	Translocation	0.000
NRAS	COSM584	p.Q61R	0.006
PIK3CA	COSM775	p.H1047R	0.006
PIK3CA	COSM12464 ²	p.N1068fs*4	0.006
EML4-ALK	NA	Translocation	0.001
ALK	COSM144250	p.G1202R	0.000
ALK	COSM28055	p.F1174L	0.000
BRCA1	COSM1383519	p.K654fs*47	0.001
BRCA2	COSM1738242	p.R2645fs*3	0.000
EGFR	COSM12370	p.L747_P753>S	0.000
EGFR	COSM6256	p.S752_I759delSPKANKEI	0.001
EGFR	COSM6223	p.E746_A750delELREA	0.008
KRAS	COSM516	p.G12C	0.001
CD74/ROS1	NA	Translocation	0.000
KRAS	COSM554	p.Q61H	0.004

N/A = not applicable

¹Variant allele frequencies > 0.00% for this wild-type negative control are within the expected range for stochastic positive dPCR reactions.

²As of June 2019, this mutation is no longer listed in the COSMIC database.

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material WT
Material Number: 0710-0674 Batch Number: 10593197
Next Generation Sequencing testing using Archer® Reveal ctDNA™ 28 Kit run on an Illumina® MiSeq™ using v2 (2x150 bp) PE chemistry reagents¹

NGS Parameters:

DNA input = 50 ng
of samples / flow cell = 1
of total reads / sample = 9.58M
Average read depth = 9064X
On-target reads = 95.9%
Q30 score = 93.3%
Analysis = Archer Analysis Suite v6.2.7 (default settings except for: Variant Downstream ROI Size set to 150 and Read Depth Normalization set to 10,000,000)

¹Please see the poster from NIST for more information about assay sensitivity:
<https://digital.seracare.com/multilab-assessment-reference-materials-ctdna-poster2018>

NOTE: Copy numbers of ERBB2 and MET were assayed by dPCR and found to be normal. They were not assayed by NGS as the wild-type sample is used as a normal control for determining copy number of genes in other samples.

Approval:

A handwritten signature in black ink, appearing to be "ERL".

Prepared By

30 NOV 2021

Date