

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseg® FFPE HRD Low-Positive Reference Material

Material Number: 0710-2645 Batch Number: 10641622

Mixture of human breast cancer cells and matched normal B

lymphoblast cells (mixed at 65% tumor) fixed with 10% formalin and Material Description:

embedded in paraffin.

Date of Manufacture: 28 OCT 2022 Expiration Date: 28 OCT 2024

2 - 8°C Storage:

Fill Volume: One 10 µm curl

DNA was extracted using the Qiagen QIAamp DNA FFPE Tissue Kit Test Method for Yield per

and quantified using the Qubit dsDNA HS Assay. Curl:

Average DNA yield per curl: 235.2 ng

Test Method for DNA **Quality Assessment:**

Agilent TapeStation Genomic DNA ScreenTape Analysis

Average DIN value: 5.6 for QIAamp extracted DNA

Illumina TruSight[™] Oncology 500 HRD RUO Assay which HRD Status Test Method:

calculates a GIS using an algorithm licensed from Myriad Genetics

Measured GIScore: 54 ± 2

Approval:

15 NOV 2022

Prepared By Date

Page 1 of 1 TPR-0710-2645 Rev. 1